

## INSTRUCTION FOR USE

### ELISA for detection of Cry2A

Cat No. AID 008



## 1 INTENDED USE

Cry2A ELISA test kit is intended to be used for the qualitative detection of Cry2A toxin in Bollgard \*\* II individual cotton leaf or seed samples. The kit is also used for detection of individual corn leaf or seed sample. The total incubation time of the assay is 60 minutes.



## 2 PRINCIPLE OF THE TEST

An antibody specific to the Cry2A protein molecule is immobilized on the microwell plates and second antibody specific for Cry2A molecule is conjugated with horse radish peroxidase (HRP). When the sample extracts are added to the microtiter wells, the Cry2A from sample binds to the antibody immobilized in the well. This binding is subsequently detected by addition of enzyme labeled antibody. After a washing step, substrate is added. The enzymatic reaction and development of color is proportional to the amount of Cry2A present in the sample. The reaction is terminated by addition of stopping solution. Absorbance is then measured on a plate reader.

Light color indicates lower concentration while dark color indicates higher concentration of Cry2A protein in the sample.

## 3 CROSS-REACTIVITY

The Cry2A ELISA kit does not recognize CP4EPSPS (Roundup Ready\*\*), Cry1F, Cry1Ac, Cry9C or PAT (Liberty Link\*\*\*).

## 4 CONTENTS OF THE KIT:

### Kits are provided in 1 plate format.

- One plate of 96 wells coated with Anti-Cry2A antibody, individually packed in a laminate bag with silica gel
- Cry2A enzyme conjugate: One Ready- to- use bottle of 6 ml
- Extraction buffer concentrate: (10X): 50 ml, 1 vial.
- Wash solution (10X): 50 ml, 1 vial.
- TMB Substrate: One ready-to-use bottle of 12 ml
- Cry2A Negative control: One ready-to-use vial of 2 ml

### Kits are provided in 10 plates format.

- Ten plates of 96 wells coated with Anti-Cry2A antibody, individually packed in a laminate bag with silica gel
- Cry2A enzyme conjugate: One Ready- to- use bottle of 55 ml
- Extraction buffer concentrate: One packet of powder with a 25 ml vial of Tween-20
- Wash solution concentrate: One packet of powder with a 2.5 ml vial of Tween-20
- TMB Substrate: One ready-to-use bottle of 110 ml

### Kits are provided in 50 plates format.

- Fifty plates of 96 wells coated with Anti-Cry2A antibody, each packed individually in a laminate bag with silica gel
- Cry2A enzyme conjugate: One ready to use bottle with 275 ml.
- Extraction buffer concentrate: One packet of powder with a 25 ml vial of Tween-20
- Wash solution concentrate: One packet of powder with a 2.5 ml vial of Tween-20
- TMB Substrate: One ready-to-use bottle of 500 ml

## ELISA for detection of Cry2A

Cat No. AID 008



- Cry2A Positive control: One ready-to-use vial of 2 ml
- Stop Solution: One ready to use vial of 12 ml.
- Cry2A Positive control: One ready-to-use vial of 10 ml
- Cry2A Negative control: One ready-to-use vial of 10 ml
- Cry2A Positive control: One ready-to-use vial of 10 ml
- Cry2A Negative control: One ready-to-use vial of 10 ml

## 5 MATERIAL AND EQUIPMENT REQUIRED BUT NOT PROVIDED

- Pipette with disposable plastic tips
- Multichannel pipette with disposable pipette tips
- Plate ELISA reader with 450/620 nm filter
- Table top centrifuge
- Marking pen, Parafilm, Pipette tips, Timer and Paper towels
- Deionized or distilled water
- Graduated cylinders with one liter capacity
- Reagent troughs
- Plate shaker
- Plate ELISA washer or wash bottle

## 6 PRECAUTIONS

The Cry2A ELISA test is intended for in vitro use only. The reagents contain Thimerosal as preservative. Prevent direct skin and eye contact with kit components. Seek medical attention in case of accidental ingestion of kit components.

## 7 STORAGE OF THE KIT

1. The kit should be stored at 2 - 8° C. The unopened kit is stable till the expiry date printed on the kit label.
2. The expiry date of each unopened component is printed on the label of the component.
3. Bring all the components of the Kit to room temperature prior to use.

## 8 REAGENT PREPARATION

### 1. Preparation of Working extraction buffer:

- Dissolve the packet of extraction buffer concentrate powder in 500 ml of distilled/deionised water and add 25 ml of Tween-20 which is provided along with the powder. Mix it well before use. The resultant extraction solution is 10X concentrate.
- This has to be further diluted with 4500 ml of water to have a working extraction buffer.
- Store unused portion of working extraction buffer at 2 - 8° C. Thaw to room temperature before using the same again.

### 2. Preparation of Working wash buffer:

- Dissolve the packet of wash buffer concentrate powder in 500 ml of distilled/deionised water and add 2.5 ml of Tween-20 which is provided along with the powder. Mix it well before use. The resultant wash solution is 10X concentrate.
- This needs to be further diluted with 4500 ml of water to have working wash buffer
- Store unused portion of working wash buffer at 2 - 8° C. Thaw to room temperature before using the same again.

## ELISA for detection of Cry2A

Cat No. AID 008



### 3. Preparation of stop solution:

- Add 27.0 ml of 98 % H<sub>2</sub>SO<sub>4</sub> in 973 ml of deionized or distilled water. Work in fume hood while preparing stop solution. Store at room temperature for up to one year.

### 4. Preparation of user positive control:

- The user is advised to prepare in-house positive control. Crush or grind one known positive seed into a uniform powder and add 1.0 ml of extraction buffer. Wait for 10 minutes before use. Prepare fresh before use.

## 9 SAMPLE PREPARATION

### 1. Individual seed

- Crush or grind seed thoroughly into a uniform powder using mortar and grinder or seed crusher. Add 1.0ml of extraction buffer to each sample. Allow the mixture to stand for 30-60 minutes at room temperature (RT). Allow particles to settle and use only the supernatant to do the test. Wash and rinse grinding equipment carefully between samples to avoid cross contamination.

### 2. Individual leaf

- Weigh 20 mg (about two leaf punch) and put it in a microtube. Crush leaf with pestle in 0.5 ml working extraction buffer. Mix well and incubate for 30-60 minutes at RT. Allow particles to settle and use only the supernatant to do the test. Avoid cross contamination between samples.

Avoid cross contamination between samples.

## 10 ASSAY PROCEDURE

Test protocol for individual seed or leaf samples

1. Allow all reagents to reach room temperature before use.
2. Add 50 µl enzyme conjugate to each well.  
Add 50 µl working extraction buffer to blank well.  
Add 50 µl of positive control in two wells and 50 µl of negative control in two wells. Add 50 µl of sample extracts to wells. Mix contents of the plate carefully, to avoid cross contamination.
3. Cover the plate and incubate it for 45 minutes at room temperature.
4. Remove content of the wells by decanting into a sink or a waste container. Add 300 µl/well wash solution to all wells and then empty wells by inverting the plate. Repeat washing three more times. Alternately, perform four washes by using microtiter plate washer. After last wash, tap the inverted microtiter plate on paper towel to remove as much liquid as possible.
5. Add 100 µl substrate per well.
6. Cover the plate and incubate for 15 minutes at room temperature.
7. Stop the reaction by adding 100 µl stop solution.
8. Measure the absorbance at 450 nm (primary filter) and 620/630 nm (secondary filter). Read the plate within 15 minutes after addition of stop solution.

Notes on Technique:

1. Protect the plates from draught, strong light or direct sunlight during the test procedure.
2. Careful aspiration of the washing solution is essential for good assay precision.
3. Since the timing of the incubation steps is important to performance of the assay, use multi channel pipettes to dispense samples/reagents.
4. Plate readers measure absorbance vertically. Do not touch the bottom of the wells.

## 11 INTERPRETATION OF RESULTS

General test criteria:

1. Absorbance of the blank well should not exceed 0.1.
2. Absorbance value of blank should be subtracted from positive control, negative control and samples.
3. The coefficient of variance between the duplicate positive control should be less than 15 %
4. Mean positive control should have absorbance of at least 0.5.
5. Mean negative control should have absorbance below 0.1.
6. If above criteria are not met, the test is invalid and should be repeated.

### Qualitative Test

1. For single corn/cotton leaf and seed samples:  
**CUTOFF CALCULATION:** for single corn/cotton leaf or seed sample Cutoff value: (Mean absorbance of negative control + 0.1)
2. **Positive sample:** If sample absorbance is more than the cutoff value.
3. **Negative sample:** If sample absorbance is less than the cutoff value.
4. Low level results normally indicate improper washing, cross contamination, or technique. In such cases, retesting is recommended.
5. Performance characteristics  
Sensitivity: 100 %  
Specificity: 100 %

## 12 WARRANTY

Gold Standard Diagnostics Hyderabad warrants that the products sold hereunder (“the Product”) are defect- free in material and workmanship, provided they are used in accordance with the prescribed instructions before the expiry of the products as printed on the product label. The customer should notify Gold Standard Diagnostics Hyderabad in writing of Warranty defects during the warranty period, including an offer by the customer to return the Products to Gold Standard Diagnostics Hyderabad for evaluation. Gold Standard Diagnostics Hyderabad will repair or replace, at its sole option, any product or part thereof that proves defective in materials or workmanship within the warranty period.

This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Gold Standard Diagnostics Hyderabad.

## 13 THIS WARRANTY IS EXCLUSIVE

The sole and exclusive obligation of Gold Standard Diagnostics Hyderabad shall be to repair or replace the defective Products in the manner and for the period provided above. Gold Standard Diagnostics Hyderabad shall not have any other obligation or liability, whatsoever it may be, with respect to the Products or any part thereof. Under no circumstances, whatsoever the circumstances may be, shall Gold Standard Diagnostics Hyderabad be liable for incidental, special, or consequential damages. If any part of this Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

\*Parafilm is a registered trademark of American Can Corporation (now Pechinney Plastic Packaging).

\*\*Bollgard & Roundup Ready are registered trademarks of the Monsanto Company.

\*\*\*Liberty Link is a register trademark of Bayer Crop Science.

**ELISA for detection of Cry2A**

Cat No. AID 008



## TECHNICAL SUPPORT SERVICE

For technical assistance and more information please contact the Eurofins Amar Immunodiagnosics Pvt. Ltd. Customer Service or your local distributor.

#1504, 15<sup>th</sup> Floor, Manjeera Trinity Corporate, eSeva Lane,  
KPHB Phase 3, Kukatpally, Hyderabad, Telangana – 500072, India.  
Tel: +91 7337480355  
E-mail: [technologies.amar@eurofins.com](mailto:technologies.amar@eurofins.com)

[www.goldstandarddiagnostics.com](http://www.goldstandarddiagnostics.com).